










Blood Collection

ORDER OF DRAW

First	Bottle/Stopper	Additive	Notes
1		Culture Bottles	See bottle label
2		Light Blue	Citrate Tube must be filled completely. Note: When using a winged blood collection set and a coagulation (citrate) tube is the first specimen, begin by drawing another partially-filled citrate tube. This discard tube is used to fill the winged set tubing's "dead space" and helps ensure a proper blood-to-additive ratio.
3		Gold	Gel, serum Do not use gel tubes for toxicology or drug testing.
4		Red	No gel, serum
5		Green or Tan	Heparin
6		Lavender or Tan	EDTA
7		Royal Blue	EDTA
8		Gray	Sodium Fluoride (Glucose)
Tubes with Other Additives			
9		Yellow	Citrate ACD This tube is drawn last.


Courtesy and © Becton, Dickinson and Company.

MIXING CHART

BD Vacutainer™ Tube Type	Stopper Color	Number of Inversions
EDTA	Lavender	8 – 10
Citrate	Light Blue	3 – 4
SST® with gel	Red/Black	5
Serum	Red	5
Sodium Fluoride	Gray	8 – 10
Heparin	Green	8 – 10

SEPARATION

For serum (no gel) tubes only.



For other blood, serum or plasma separation instructions, see the Specimen Collection and Handling section of this directory.