

4000 Saginaw St, Flint, MI 48507

Specimen Collection Tubes

BLOOD SPECIMENS		Always refer to tube label to confirm tube type - never rely on the stopper color alone.
Stoppe	r/Label Color	Laboratory Use Additives/Inversions at Collection
1	Culture Bottles	For detection of microbial growth from blood specimens. • Soybean-casein digest broth 8-10 gentle inversions unless otherwise noted
2	Light Blue	For coagulation determinations. Note: Certain tests may require chilled/refrigerated specimens. Follow recommended procedures for collection and transport. Inversions prevent clotting. • 0.105 M sodium citrate (3.2%) 3-4 inversions
3	Gold	Serum Separator Tube (SST®) for serum determinations in chemistry and serology. Contains separator gel and should not be used for toxicology or drug testing. Inversions ensure mixing of clot activator with blood. Blood clotting time 30 minutes. • Clot activator and gel for serum separation • Serum Separator Tube (SST®) for serum determinations in serum separation • Clot activator and gel for serum separation • Serum Separator Tube (SST®) for serum determinations in serum separation • Clot activator and gel for serum separation • Serum Separator Tube (SST®) for serum determinations in serum separation • Clot activator and gel for serum separation • Serum Separator Tube (SST®) for serum determinations in serum separation • Serum Separator Tube (SST®) for serum determinations in serum separation • Serum Separator Tube (SST®) for serum determinations in serum separation • Serum Separator Tube (SST®) for serum determinations in serum separation • Serum Separator Tube (SST®) for serum determinations in serum separation • Serum Separator Tube (SST®) for serum determinations in serum separation • Serum Separator Tube (SST®) for serum determinations in serum separation • Serum Separator Tube (SST®) for serum determinations in serum separation • Serum Separator Tube (SST®) for serum separation serum separation serum
4	Red	For serum determinations in chemistry and serology, and for toxicology and drug testing. Glass serum tubes are recommended for blood banking. Plastic tubes contain clot activator and are not recommended for blood banking. Inversions ensure mixing of clot activator with blood and clotting within 30-60 minutes.
5/6	Green	For plasma determinations in chemistry. Inversions prevent clotting. Use only sodium heparin green-top tubes for all cytogenetic testing. Sodium heparin Lithium heparin 8-10 gentle inversions
5/6	Tan	For lead determinations. This tube is certified to contain less than 0.01 µg/mL (ppm) lead. Inversions prevent clotting. • K ₂ EDTA (plastic) 8-10 gentle inversions
5/6	Lavender	K ₂ EDTA for whole-blood hematology determinations and immunohematology testing (ABO grouping, Rh typing, antibody screening). Inversions prevent clotting.
7	Royal Blue	 K2 EDTA for whole-blood hematology determinations and immunohematology testing (ABO grouping, Rh typing, antibody screening). Inversions prevent clotting. • Spray-dried K ₂ EDTA 8-10 gentle inversions
8	Gray	For glucose determinations. Oxalate and EDTA anticoagulants will give plasma samples. Sodium fluoride is the antiglycolytic agent. Inversions ensure proper mixing of additive and blood. • Potassium oxalate/sodium fluoride and sodium fluoride oxalate/sodium fluoride.
9	Yellow and white label	Glass tube with liquid ACD for use in blood bank studies, HLA phenotyping, DNA, paternity testing, etc. - Acid Citrate Dextrose (ACD) solutions A/B additives – Trisodium citrate 22.0/13.2, citric acid 8.0/4.8 and dextrose 24.5/14.7 (in g/L) 8-10 gentle inversions
	Note: The QuantiFERON®-To	3 Gold collection tube set includes lavender, gray and purple capped tubes that are not listed or indicated on this chart
URIN	ESPECIMENS	
Stoppe	r/Label Color	Laboratory Use Additives/Inversions at Collection
	Gray and yellow label	For culture and sensitivity (C&S) urine testing. Minimum urine volume is 4 mL. For lower volumes, submit refrigerated urine in a sterile container without preservatives. • Boric acid, sodium formate Shake vigorously

For questions regarding specimen collection tubes not shown here, contact Client Services or your Sales Representative.

For urinalysis testing. Inversions ensure preservative is

or over fill (>10 mL).

properly mixed. Note the fill lines. Do not under fill (<2 mL)

Yellow plastic

and yellow label

Preservative
 8-10 gentle inversions